ELSEVIER

Journal of Chromatography A, 708 (1995) 99–104

JOURNAL OF
CHROMATOGRAPHY A

Rapid determination of germanium and tin by ion chromatography

Qun Sun, Huitong Wang, Shifen Mou*

Research Centre for Eco-Environmental Sciences, Academia Sinica, P.O. Box 2871, Beijing 100085, China

First received 17 June 1993; revised manuscript received 14 March 1995; accepted 17 March 1995

Abstract

A method was established for the simultaneous determination of germanium and tin in one run by ion chromatography. Using a selected eluent, Ge and Sn were separated on a bifunctional ion-exchange column, followed by a postcolumn chromogenic reaction and spectrophotometric detection at 520 nm. The whole procedure was completed within 5 min with high sensitivity and selectivity at the low- $\mu\text{g/l}$ level. The results of real sample analysis were satisfactory.

1. Introduction

In research in the metallurgical industry and on the analysis of mineral resources, the determination of Ge and Sn is necessary. During the past 30 years, with the discovery of large amounts of germanium in many kinds of rare medicinal herbs, the function of germanium in pharmacology and biology has attracted increasing interest. It has been found that germanium is a very important element in the human body which can improve the physical condition [1]. Current methods for the determination of Ge and Sn include spectrophotometric measurement [2–8], atomic absorption spectrometry (AAS) [9–11] and polarography [12,13]. Among these methods, spectrophotometry has frequently been used but it is less sensitive and selective and always needs different sample preparation. AAS with hydride generation is a relatively powerful

way to determine Ge and Sn, with detection limits of 0.6 ng/l for Ge and 1.7 ng/l for Sn [1,9,14]. However, ion chromatography has not been reported for the simultaneous determination of Ge and Sn. In this paper, we present a rapid method of ion chromatographic separation followed by spectrophotometric detection which shows high sensitivity and a lack of interferences.

2. Experimental

2.1. Materials

A stock standard solution of germanium (1000 mg/l) was obtained by dissolving germanium powder in 30% hydrogen peroxide solution with solid sodium hydroxide [8]. The pH of the Ge solution was adjusted to <2 with sulfuric acid.

A stock standard solution of Sn (1000 mg/l) was prepared by dissolving tin metal in hot, concentrated hydrochloric acid and adding hy-

* Corresponding author.

droxylamine hydrochloride to reduce Sn(IV) to Sn(II).

The eluent was a mixture of 0.15 mol/l sulfuric acid and 0.05 mol/l sodium sulfate.

A phenylfluorone stock standard solution (1.5 mmol/l) was prepared by dissolving 0.48 g of phenylfluorone in 500 ml of ethanol, adding 50 ml of 6.0 mol/l sulfuric acid and filtering through a 0.45- μ m filter-paper. The filtrate was diluted to 1000 ml with deionized water.

A 2% solution of the cationic surface-active agent cetyltrimethylammonium bromide (CTAB) was prepared by dissolving 10 g of CTAB in 400 ml of hot water and diluting to 500 ml with deionized water.

A 0.20 mol/l oxalic acid solution was prepared by dissolving 16.2 g of oxalic acid in 500 ml of deionized water.

The postcolumn reagent was prepared by mixing 250 ml of 1.5 mmol/l phenylfluorone with 120 ml of 2% CTAB and 50 ml of 0.2 mol/l oxalic acid, adding 40 ml of concentrated sulfuric acid and diluting to 500 ml with deionized water.

All chemicals were of analytical-reagent grade from China.

2.2. Chromatographic apparatus

All chromatography was performed on a Dionex 2010 IC system equipped with a basic ion chromatographic module, a spectrophotometric detector, a postcolumn reactor and a 50- μ l sample loop. The separation column was a Dionex IonPac CS5 with an IonPac CG5 guard column. The chromatogram was plotted by a Dionex recorder.

2.3. Procedure

Fig. 1 shows the chromatographic configuration for the determination of Ge and Sn. The pump delivered the H_2SO_4 – Na_2SO_4 eluent at the rate of 1.0 ml/min. The postcolumn reactor provided a flow-rate of 0.5 ml/min for the above postcolumn reagent. The wavelength for determination of Ge and Sn was set at 520 nm. In order to make the postcolumn chromogenic reaction complete, the eluent and postcolumn

Fig. 1. Chromatographic configuration.

reagent must be mixed in a beaded mixing coil 80 cm long. Fig. 2 shows the chromatogram of germanium and tin.

Fig. 2. Chromatogram of germanium and tin. Peaks: 1 = HGeO_3^- (0.5 mg/l); 2 = Sn^{2+} (1.0 mg/l). Separation column, Dionex IonPac CS5; guard column, IonPac CG5; sample loop, 50 μ l; detection wavelength, 520 nm; eluant, 0.15 mol/l H_2SO_4 –0.05 mol/l Na_2SO_4 ; flow-rate, 1.0 ml/min; postcolumn reagent, 0.75 mmol/l phenylfluorone–0.5% CTAB–1.5 mol/l H_2SO_4 –0.02 mol/l oxalic acid; flow-rate, 0.5 ml/min.

3. Results and discussion

3.1. Choice of separation column

The stable oxidation state of germanium in solution is Ge(IV) [1]. In aqueous acidic solution, Ge exists primarily in its oxy acid anion form, HGeO_3^- . The stable valency of tin in solution is Sn(II) [14]. In order to separate HGeO_3^- and Sn^{2+} simultaneously, we chose a Dionex IonPac CS5 column as the separation column, as its resin contains both anion- and cation-exchange groups.

3.2. Selection of eluent

In acidic solution, HGeO_3^- is stable but Sn^{2+} can be hydrolyzed to SnOH^+ at $\text{pH} > 1.5$. The eluent used for separating HGeO_3^- and Sn^{2+} should be maintained at $\text{pH} < 1.5$ to keep them in their stable state. In order to separate HGeO_3^- anions and Sn^{2+} cations simultaneously in one run, the eluent should contain both anions and cations since the separation mechanism of the IonPac CS5 column is ion exchange.

Initially, HCl was used as the eluent, but more than two peaks were observed and the reproducibility of the response was very poor. The reason could be the complex reaction between germanium and HCl. Germanium can form GeCl_4 , GeCl_5^- , GeCl_6^{2-} and H_2GeO_3 in HCl [1]. Owing to the presence of H_2SO_4 in the post-column chromogenic reagent, H_2SO_4 was selected for a trial. However, the results showed that the retention time of Sn^{2+} was too long with elution using H_2SO_4 , as H^+ is a relatively weak eluting ion which has a weak affinity to the cation-exchange resin of the IonPac CS5 column. To decrease the retention time of Sn^{2+} and maintain the retention time of HGeO_3^- , we added Na_2SO_4 to the H_2SO_4 eluent, as Na^+ is a stronger eluting ion than H^+ . When sodium sulfate was added to the eluent, the concentration of sulfuric acid must be lowered, as the sodium sulfate contributes not only Na^+ cations but also SO_4^{2-} anions to the eluent. While keeping the concentration of SO_4^{2-} constant, we

Table 1

Effect of concentrations of H_2SO_4 and Na_2SO_4 on the retention times of HGeO_3^- and Sn^{2+}

Na_2SO_4 (mol/l)	H_2SO_4 (mol/l)	Retention time (min)	
		HGeO_3^-	Sn^{2+}
0	0.20	2.0	6.2
0.05	0.15	2.0	4.0
0.10	0.10	2.0	3.2
0.20	0	2.1	2.0

changed the $\text{H}_2\text{SO}_4/\text{Na}_2\text{SO}_4$ concentration ratio in the eluent. Table 1 gives the results of the effect of the concentration of H_2SO_4 and Na_2SO_4 on the resolution of Ge and Sn. With increase in the concentration of Na^+ , the retention time of Sn^{2+} decreased, whereas the retention time of HGeO_3^- remained the same. We adopted 0.15 mol/l H_2SO_4 –0.05 mol/l Na_2SO_4 as the eluent.

3.3. Selection of chromogenic reagent

Since the late 1960s, phenylfluorone and CTAB have been used as effective chromogenic reagents in spectrophotometry for the determination of Ge(IV) and Sn(II) [7]. However, the postcolumn reaction in ion chromatography is different from that in spectrophotometry; it occurs under fluxional conditions within a short reaction time. Hence the chromogenic conditions of spectrophotometry cannot be used directly, and should be modified. In order to make the postcolumn reaction complete and rapid, the conditions for ion chromatography were studied as discussed below.

Concentration of phenylfluorone

Phenylfluorone (9-phenyl-2,3,7-trihydroxy-6-fluorone) is a highly sensitive chromogenic reagent for Ge(IV) and Sn(II). The peak heights of Ge and Sn were in direct proportion to the concentration of phenylfluorone. However, when the concentration of phenylfluorone was > 0.5 mmol/l, the peak heights of Ge and Sn became constant. Therefore, 0.75 mmol/l

phenylfluorone in the postcolumn reagent was selected for the determination of Ge and Sn.

Effect of concentration of H_2SO_4

The function of H_2SO_4 in the postcolumn reaction is to adjust the acidity of the reagent, which directly affects the absorbance of the complexes of $HGeO_3^-$ and Sn^{2+} . The experiment showed that, with increase in the concentration of H_2SO_4 , the response of $HGeO_3^-$ and Sn^{2+} was improved until the concentration of H_2SO_4 reached 1.0 mol/l. At concentrations of H_2SO_4 between 1.0 and 2.0 mol/l, the absorbances of $HGeO_3^-$ and Sn^{2+} were constant. Hence 1.5 mol/l H_2SO_4 was used in the postcolumn reaction.

Concentration of CTAB

The cationic surface-active agent CTAB plays a very important role in the postcolumn reaction. It forms tri-complex compounds with $HGeO_3^-$ or Sn^{2-} and phenylfluorone, which increase the absorbance of the products in the postcolumn chromogenic reaction [7,8]. Also, it is a solubilizing agent in the reaction. The experiments showed that the optimum concentration of CTAB for the postcolumn reaction is 0.5% (v/v).

Effect of oxalic acid

Oxalic acid can catalyse the chromogenic reaction of phenylfluorone with Sn^{2+} , but it does not have any effect on the $HGeO_3^-$ reaction [15]. When oxalic acid was added to the postcolumn reagent, the peak height of Sn^{2+} in the chromatogram increased sharply whereas that of $HGeO_3^-$ remained the same. When the concentration of oxalic acid was >0.016 mol/l, the peak height of Sn^{2+} became constant. It was also found that the interference in the determination of Ge in the presence of large amounts of Sn^{2+} could be easily removed by changing the concentration of oxalic acid. For the simultaneous determination of germanium and tin, 0.02 mol/l oxalic acid in the postcolumn reagent was adopted.

Based on the above discussion, the optimum composition of the postcolumn reagent is 0.75

mmol/l phenylfluorone, 0.5% CTAB, 1.5 mol/l H_2SO_4 and 0.02 mol/l oxalic acid.

3.4. Interferences

It has been reported that some metal ions also react with phenylfluorone [5,6,8]. However, in this method, with elution by H_2SO_4 – Na_2SO_4 , low-valent metal ions (such as alkali and alkaline earth metals and other divalent ions, Ni^{2+} , Fe^{2+} , Cu^{2+} , Co^{2+} , Zn^{2+} , etc.) cannot be retained on the IonPac CS5 separation column and are rapidly removed with the system peak; however, high-valent metal ions such as Fe^{3+} , Al^{3+} and Mo(VI) were retained on the column for a long time. The experiments showed that there was no interference in the determination of Ge and Sn when Fe^{3+} , Al^{3+} , and Mo(VI) co-existed in the samples (when their concentration was <1000 times greater than that of $HGeO_3^-$ and <500 times greater than that of Sn^{2+}). However, oxidants such as perchloric acid interfere in the determination of Sn^{2+} . Compared with other methods for the determination of germanium and tin, the proposed method is simple, sensitive and highly selective and the interferences can be easily removed.

3.5. Detection limits, linearity and reproducibility

Table 2 gives the results for linearity, reproducibility and sensitivity. The results are satisfactory.

3.6. Results for real samples

We applied the method to the determination of $HGeO_3^-$ and Sn^{2-} in two kinds of samples: an alloy and an organic germanium nutritive liquid. Fig. 3 shows the chromatograms.

The preparation for the alloy involved dissolving 1.0 g of alloy in aqua regia, adding hydroxylamine hydrochloride in order to reduce Sn(IV) to Sn(II) and diluting to 100 ml with deionized water.

The organic germanium nutritive liquid is a newly developed functional health product

Table 2

Linearity, detection limits and relative standard deviations for the determination of HGeO_3^- and Sn^{2+}

Ion	Linearity, r	R.S.D. (%) ($n = 7$)	Detection limit ($\mu\text{g/l}$) ^a
HGeO_3^-	0.9998 (0.5–1000 $\mu\text{g/l}$)	1.8 (0.05 mg/l)	0.5
Sn^{2+}	0.9999 (1.0–30.0 mg/l)	1.1 (15.0 mg/l)	1.5
	0.9990 (1.5–2000 $\mu\text{g/l}$)	2.0 (0.1 mg/l)	

^a Signal-to-noise ratio = 3. Fig. 3. Chromatograms for real sample analysis: (A) alloy; (B) organic germanium nutritive liquid. Peaks: 1 = HGeO_3^- ; 2 = Sn^{2+} .

(Tianlong Health Nutriment, China). The sample was prepared by mixing 1 ml of the liquid with 5 ml of concentrated nitric acid and 1 ml

hydrogen peroxide, digesting for 2 h and diluting to 100 ml with deionized water.

Spike and recovery studies were performed using the above samples and AAS was also applied for comparison purposes. The results are given in Table 3. Good recoveries for the determination of Ge and Sn were obtained (96–102%), and the results with the proposed method were in good agreement with the AAS data.

References

- [1] K. Wang, Trace Elements in Life Science, China Measurement Publishing House, Beijing, 1992, p. 511.
- [2] C.L. Leong, Talanta, 18 (1971) 845.
- [3] H.B. Corbin, Anal. Chem., 45 (1973) 534.
- [4] Y. Chen and B. Wang, Fenxi Huaxue, 17 (1989) 498.
- [5] A. Harada and T. Tarutani, Anal. Chim. Acta, 209 (1988) 333.
- [6] H. Shen, Z. Wang and G. Xu, Analyst, 112 (1987) 887.
- [7] Y. Shijo and T. Takeuchi, Bunseki Kagaku, 51 (1967) 161.

Table 3

Results for real samples

Sample	Ion	Found ^a (mg/l)	Added (mg/l)	Result (mg/l)	Recovery (%) ^b	Total content ^c (mg/l)
Alloy (1 g)	HGeO_3^-	0.0	0.25	0.24	96	—
	Sn^{2+}	0.46	0.5	0.97	102	46
Ge liquid (1 ml)	HGeO_3^-	15.82	9.9	25.42	97	1582
	Sn^{2+}	0.0	0.25	0.25	100	—

^a Found in analysed samples, i.e., after dilution.^b Recovery = (results – found) × 100/addition.^c Total content (mg/l) of HGeO_3^- and Sn^{2+} in the original (undiluted) samples (1 g of alloy or 1 ml of Ge liquid).^d AAS data.^e Approximate value given by the producer.

- [8] D. Shang and H. Xu, *Fenxi Huaxue*, 3 (1975) 203.
- [9] M.O. Andreae and P.N. Froelich, Jr., *Anal. Chem.*, 53 (1981) 287.
- [10] L. Halicz, *Analyst*, 110 (1985) 943.
- [11] J.R. Castillo, *At. Spectrosc.*, 2 (1981) 159.
- [12] C.M. Joy, *Anal. Chim. Acta*, 142 (1982) 231.
- [13] N. Li, *Huaxue Xuebao*, 45 (1987) 399.
- [14] K. Wang, *Trace Elements in Life Science*, China Measurement Publishing House, Beijing, 1992, p. 540.
- [15] H. Xu, D. Zhang and X. Lin, *Fenxi Huaxue*, 4 (1976) 218.